



DOC226187 F Valid from Jan 1st 2021

DL1416 Temperature Logger

Please choose one attribute from each numbered category and add the corresponding mark to the grid on top of the table.

As a result configuration code for the Vaisala device is formed. Example of order code with typical settings:

DL1416 V A 1 A 1 N N N N N

Selections in bold are included in the prices of the basic versions.

Selections in italic are available at an extra price.

| Vaisala Data Logger Configuration | | DL1416 | | | | | | | N | Ν | 1 |
|-----------------------------------|-----------------------------------|----------------------|----------------|------------|---|---|---|---|---|---|---|
| 1. Validation Type | Validation | | V | | | | Г | | | | Ī |
| | Non Validation | | S | | | 1 | | | | | ı |
| 2. Channel Function | Temperature Stainless Steel | 7.6 m (25 ft.) Probe | — _/ | ۱ | | | | | | | ı |
| Length Ext 1 | Temperature Water Proof | 7.6 m (25 ft.) Probe | Ε | 3 | | 1 | | | | | ı |
| 3. Channel Function | Temperature Stainless Steel | 7.6 m (25 ft.) Probe | | - 1 | | | | | | | ı |
| Length Ext 2 | Temperature Water Proof | 7.6 m (25 ft.) Probe | | 2 | | | | | | | ı |
| | Boolean | 7.6 m (25 ft.) | | 3 | | | | | | | ı |
| | Boolean | No cables | | 4 | | | | | | | ı |
| 4. Channel Function | Temperature Stainless Steel | 7.6 m (25 ft.) Probe | | | A | | | | | | ı |
| Length Ext 3 | Temperature Water Proof | 7.6 m (25 ft.) Probe | | | В | | | | | | ı |
| 5. Channel Function | Not used | | | | | N | | | | | ı |
| Length Ext 4 | Temperature Stainless Steel | 7.6 m (25 ft.) Probe | | | | 1 | | | | | ı |
| | Temperature Water Proof | 7.6 m (25 ft.) Probe | | | | 2 | | | | | ı |
| | Boolean | 7.6 m (25 ft.) | | | | 3 | | | | | ı |
| | Boolean | No cables | | | | 4 | | | | | ı |
| 6. Additional Temperature | None | | | | | | N | | | | ı |
| Calibration Points* | 1 Additional Point** Specify Add | itional Point here | | | | | Α | | | | ı |
| | 2 Additional Points** Specify Add | itional Points here | | | | | В | | | | ı |
| 7. Preset Calibration | Recommended | 12 Months | | | | | | N | | | ı |
| Notification Interval | Customer Specified | 3 - 120 Months | | | | | | 1 | | | ı |

^{*}Additional 1 week delivery

| **Additional Calibration Points | Temperature Range | | | | | | |
|--|-------------------|--|--|--|--|--|--|
| Stainless Steel and Water Proof Probes | -90 to +50 Deg C | | | | | | |