

Certificate Number: AHEL192050009



Case Number: 123456
Customer: Vaisala Oyj
Address: Vanha Nurmijarventie 21
01670 Vantaa
Finland

Manufacturer: Vaisala Oyj
Instrument: Carbon Dioxide Probe GMP251
Serial Number: P2130092
Calibration Date: 2019-05-16
Recalibration Date: 2020-05-16

Approved by:

Aki Jurvanen
Senior Test Engineer

Note(s):
Service report as an attachment.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95 %. The measurement results are traceable to the international system of units (SI) through national metrology institutes (NIST USA or equivalent) or via ISO/IEC 17025 accredited calibration laboratories.

The measurements carried out and the Certificates of Calibration issued by an Accredited Calibration Laboratory comply with the measurement ranges and uncertainties approved by FINAS Finnish Accreditation Service. The measurement results issued by the Laboratory are traceable to national or international measurement standards. Measurement Standards Laboratory of Vaisala Oyj is a calibration laboratory K008 accredited by FINAS Finnish Accreditation Service, accreditation requirement ISO/IEC 17025. The accreditation is included in the Multilateral Agreement (EA MLA) of the European co-operation for Accreditation (EA).

Certificate Number: AHSL192050009



As Found Results

Carbon dioxide concentration (percent volume) calibration results

Reference [%CO ₂]	Calibration pressure [hPa]	Reading [%CO ₂]	Error [%CO ₂]	Uncertainty [%CO ₂]	Specification [%CO ₂]	Notes
0.500	1025.3	0.526	0.026	0.011	0.200	-
5.002	1025.1	5.164	0.162	0.056	0.100	-
19.718	1024.8	20.442	0.724	0.208	0.400	*

Temperature calibration results

Reference [°C]	Reading [°C]	Error [°C]	Uncertainty [°C]	Specification [°C]	Notes
23.0	24.6	1.6	0.6	-	-

Analog channel 1 calibration results

Output settings: 0 to 5 V from 0 to 20 %CO₂

Reference [%CO ₂]	Calibration pressure [hPa]	Reading [V]	Reading [%CO ₂]	Error [%CO ₂]	Uncertainty [%CO ₂]
0.500	1025.3	0.131	0.524	0.024	0.014
5.003	1025.0	1.291	5.164	0.161	0.060
19.715	1024.8	5.078	20.312	0.597	0.217

Analog channel 2 calibration results

Output settings: 4 to 20 mA from 0 to 20 %CO₂

Reference [%CO ₂]	Calibration pressure [hPa]	Reading [mA]	Reading [%CO ₂]	Error [%CO ₂]	Uncertainty [%CO ₂]
0.500	1025.3	4.415	0.519	0.019	0.014
5.003	1025.0	8.124	5.155	0.152	0.060
19.715	1024.8	20.235	20.294	0.579	0.215

References used in the calibration

Model	Instrument number	Certificate number	Calibration date	Recalibration date
PXIe-4080	18141	B03517	2018-12-04	2019-12-31
PTU307	18002	K008-B03240	2018-11-07	2019-11-30
SEC-Z512MGX	18292	C01223	2019-04-11	2020-04-30
SEC-Z512MGX	18308	C01224	2019-04-11	2020-04-30
SEC-Z512MGX	18309	C01225	2019-04-10	2020-04-30

Ambient conditions:

Humidity

23.1 %rh ±1.6 %rh

Temperature

23.0 °C ±1.2 °C

Pressure

1025.3 hPa ±1.4 hPa

Any error greater than the specification is noted with *

Certificate Number: AHEL192050009



As Left Results

Carbon dioxide concentration (percent volume) calibration results

Reference [%CO ₂]	Calibration pressure [hPa]	Reading [%CO ₂]	Error [%CO ₂]	Uncertainty [%CO ₂]	Acceptance Limit [%CO ₂]	Pass/Fail
0.500	1025.3	0.500	0.000	0.011	0.200	PASS
5.002	1025.1	5.002	0.000	0.056	0.100	PASS
19.718	1024.8	19.718	0.000	0.208	0.400	PASS

Temperature calibration results

Reference [°C]	Reading [°C]	Error [°C]	Uncertainty [°C]	Specification [°C]	Notes
23.0	23.0	0.0	0.6	-	-

Analog channel 1 calibration results

Output settings: 0 to 5 V from 0 to 20 %CO₂

Reference [%CO ₂]	Calibration pressure [hPa]	Reading [V]	Reading [%CO ₂]	Error [%CO ₂]	Uncertainty [%CO ₂]
0.500	1025.3	0.123	0.492	-0.008	0.014
5.003	1025.0	1.257	5.028	0.025	0.060
19.715	1024.8	4.926	19.704	-0.011	0.217

Analog channel 2 calibration results

Output settings: 4 to 20 mA from 0 to 20 %CO₂

Reference [%CO ₂]	Calibration pressure [hPa]	Reading [mA]	Reading [%CO ₂]	Error [%CO ₂]	Uncertainty [%CO ₂]
0.500	1025.3	4.395	0.494	-0.006	0.014
5.003	1025.0	8.020	5.025	0.022	0.060
19.715	1024.8	19.764	19.705	-0.010	0.215

References used in the calibration

Model	Instrument number	Certificate number	Calibration date	Recalibration date
PXIe-4080	18141	B03517	2018-12-04	2019-12-31
PTU307	18002	K008-B03240	2018-11-07	2019-11-30
SEC-Z512MGX	18292	C01223	2019-04-11	2020-04-30
SEC-Z512MGX	18308	C01224	2019-04-11	2020-04-30
SEC-Z512MGX	18309	C01225	2019-04-10	2020-04-30

Ambient conditions:

Humidity

23.1 %rh ±1.6 %rh

Temperature

23.0 °C ±1.2 °C

Pressure

1025.3 hPa ±1.4 hPa

Pass/Fail: Pass - Error less than Acceptance Limit, Fail - Error equal to or more than Acceptance Limit

Certificate Number: AHEL192050009



Calibration note(s):

The carbon dioxide probe was calibrated by comparing the probe's carbon dioxide concentration reading to a generated reference carbon dioxide concentration reading in a calibration adapter. The carbon dioxide concentration readings were read with resolution of 0.1 ppmv. The reference carbon dioxide concentration reading was generated by mixing pure carbon dioxide gas and purified nitrogen gas. The mixed gas flow through the calibration adapter was approximately 0.7 liters \pm 0.2 liters per minute. The pressure compensation of the probe was on during the calibration, P comp mode = On. The calibration pressure was measured before each calibration point and inputted into the memory of the probe. The temperature compensation of the probe was on and the probe was configured to use its internal temperature value as a compensation temperature value, T comp mode = Measured. O2 and RH comp modes were off. Before measurements the probe was allowed to stabilize to the conditions of the laboratory for at least 1 hour with 24 Vdc \pm 0.3 Vdc power supply on.

The temperature of the probe was calibrated by comparing the probe's temperature reading to a reference thermometer. The temperature is always adjusted when any as left results are given to achieve best carbon dioxide concentration accuracy.

The calibration results relate only to the calibrated instrument and the calibration points. The calibration results were obtained from the measured values or the results were calculated from the measured values by using adjustment coefficients. The analog channels are always adjusted when possible if any digital or analog as left results are given. Only the analog values inside the analog range of the probe are shown in the analog results.

The uncertainty of the 0 ppm point includes the difference between 0 ppm and the negative reading of the probe when the 0 ppm point was measured if the reading of the probe froze to display 0 ppm before reaching the 0 ppm point.