

Calibration Certificate

Certificate Number: K008-1026599
Customer: Vaisala Oyj
Box26
00421 Helsinki
Finland
Manufacturer: Vaisala Oyj
Instrument: VL-1400-44N Temperature Logger
Serial Number: 16111012

Page 1 (1)

The external temperature sensor(s) of the above described temperature logger were calibrated by comparing the logger readings to a reference thermometer in Vaisala Measurement Standards Laboratory (MSL).
The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95 %.
Traceable to the National Institute of Standards and Technology (NIST, USA) via MSL.

Measurement results after adjustment:

| Reference | | Channel 1 | | | | Channel 2 | | | | |
|-------------|----------|-----------|-------------|-------|-----------|-----------|-------|-------------|-------|-----------|
| Temperature | Observed | Error | Uncertainty | Limit | Pass/Fail | Observed | Error | Uncertainty | Limit | Pass/Fail |
| °C | °C | °C | °C | °C | | °C | °C | °C | °C | |
| -25.00 | -25.00 | 0.00 | 0.14 | ±0.20 | Pass | -25.00 | 0.00 | 0.14 | ±0.20 | Pass |
| 10.00 | 10.00 | 0.00 | 0.06 | ±0.20 | Pass | 10.00 | 0.00 | 0.06 | ±0.20 | Pass |
| 25.00 | 25.00 | 0.00 | 0.06 | ±0.10 | Pass | 25.00 | 0.00 | 0.06 | ±0.10 | Pass |
| 45.00 | 45.00 | 0.00 | 0.08 | ±0.20 | Pass | 45.00 | 0.00 | 0.08 | ±0.20 | Pass |
| 69.99 | 69.99 | 0.00 | 0.12 | ±0.20 | Pass | 69.99 | 0.00 | 0.12 | ±0.20 | Pass |

| Reference | | Channel 3 | | | | Channel 4 | | | | |
|-------------|----------|-----------|-------------|-------|-----------|-----------|-------|-------------|-------|-----------|
| Temperature | Observed | Error | Uncertainty | Limit | Pass/Fail | Observed | Error | Uncertainty | Limit | Pass/Fail |
| °C | °C | °C | °C | °C | | °C | °C | °C | °C | |
| -25.00 | -25.00 | 0.00 | 0.14 | ±0.20 | Pass | -25.00 | 0.00 | 0.14 | ±0.20 | Pass |
| 10.00 | 10.00 | 0.00 | 0.06 | ±0.20 | Pass | 10.00 | 0.00 | 0.06 | ±0.20 | Pass |
| 25.00 | 25.00 | 0.00 | 0.06 | ±0.10 | Pass | 25.00 | 0.00 | 0.06 | ±0.10 | Pass |
| 45.00 | 45.00 | 0.00 | 0.08 | ±0.20 | Pass | 45.00 | 0.00 | 0.08 | ±0.20 | Pass |
| 69.99 | 69.99 | 0.00 | 0.12 | ±0.20 | Pass | 69.99 | 0.00 | 0.12 | ±0.20 | Pass |

| Reference(s): | Serial Number | Calibration Date | Certificate | Next Calibration |
|-----------------|---------------|------------------|-------------|------------------|
| Fluke 1560/2560 | TP 14410/1 | 2015-12-25 | K008-Y03121 | 2016-06-30 |


Note(s):

This is new instrument without before adjustment data

Calibration Date: April 18, 2016 **Next Calibration:** April 18, 2017

Date: April 19, 2016

Signature:



Susanna Heikkilä
Calibration Technician

Digitally signed by SHEI
Date: 2016.04.19 08:44:43 +03:00
Reason: Calibration responsible
Location: Vaisala Oyj

Ambient Condition(s):

22.1 °C ±0.2 °C
33.1 %RH ±1.4 %RH
998.9 hPa ±0.4 hPa

This Certificate may only be reproduced in full, except with the prior written permission by the issuing Laboratory.
Vaisala Measurement Standards Laboratory is a calibration laboratory K008 accredited by FINAS Finnish Accreditation Service, accreditation requirement EN ISO/IEC 17025. The accreditation is included in the Multilateral Agreement (EA MLA) of the European co-operation for Accreditation (EA).

Vaisala Oyj, PO Box 26, FI-00421 Helsinki, Finland Telephone +358 9 894 91 | Fax +358 9 8949 2227
Email helpdesk@vaisala.com | www.vaisala.com Domicile Vantaa, Finland | VAT FI01244162 | Business ID 0124416-2